

C E R T I F I C A T E

Specific test report

[in accordance with DIN EN 10204]

Device Name VIB 5.310

Device Serial No. 33015

We hereby testify that this instrument and its accessories have been manufactured according to our quality standards and have been tested according to the currently applicable specifications.

The functionality, calibration, and workmanship have been found to conform to the currently valid test criteria and within applicable tolerances.

The above instrument has been tested and calibrated on 2017-12-02.

The recommended calibration interval is 24 months.

Calibration / test equipment:

Calimaster: VIB 2.251

S/N: 5

The used calibration / test equipment is traceable to national standards and calibrated regularly.

Prüftechnik Condition Monitoring GmbH
WernerKahles

Oskar-Messter-Str. 19-21
Ismaning
85737
Germany

+49 89 99616 - 0
+49 89 99616 - 300
www.pruftechnik.com


Inspector: Werner Kahles