



The testing of frequency meter belts TEN-SIT is made on 100% of production following the specifications herewith below indicated:

Tolerance	up to 400 Hz	$\leq \pm 1\%$	± 1 digit
	up to 600 Hz	$\leq \pm 2\%$	± 1 digit

Tested values: 8 Hz

8 Hz
15 Hz
25 Hz
50 Hz
100 Hz
200 Hz
400 Hz
500 Hz
600 Hz

In addition to frequency measurements the employed must also carry out:

- a) Check of welding state
- b) Check of cardholder and carrying case state

Code TEN-SIT certificate: 19-3836/2.0Date: 29/01/2019

SIT S.p.A.
Stabilimento
24012 VALBREMBILLA (BG) Via G. Garibaldi
Inspector signature: Tel. 0345/98151 - Fax 0345/98374
Cod. Fisc. 00821330163
Part. IVA 079493360155

SIT SPA guarantees the recalibration of instrument if necessary